

BIOSYSTEMS A25

CHOLESTEROL

CHOD-PAP

Liquid
Single Reagent

REF

Cont.

D96112B 1 x 1000 ml Single Reagent

D95116 5 x 100 ml Single Reagent

D95114 1 x 3 ml Cholesterol Standard

D98485	5 x 3 ml	Calibrator	Diacal Auto
D99486	3 x 3 ml	Lipid Control normal	Diacon Lipids
D98481	12 x 5 ml	Control normal	Diacon N
D98482	12 x 5 ml	Control abnormal	Diacon P

1. Reagent preparation:

The reagent is ready to use.

2. Instrument settings:

Temperature: 37 °C

GENERAL	Test name	Choleterol
	Analysis mode	Endpoint
	Sample type	Serum
	Units	mg/dL
	Reaction type	Ascending
	Decimals	0
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Bichromatic
Volumes	Sample	3
	Reagent 1	300
	Reagent 2	-
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	505
	Reference	670
Times	Reading 1	600
	Reading 2	-
	Reagent 2	-
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	3
	Calibration curve	-
OPTIONS	Blank absorbance limit	0,20
	Kinetic blank limit	-
	Linearity limit	800

NOTE:

These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.

:

-