APPLICATION PROCEDURE

Biosystems A25

CREATININE

"Mod. Jaffe"

Liquid
2 Reagents

REF Cont.

D95595	**5 x 100 ml**	4 x 100 ml	Reagent 1
		1 x 100 ml	Reagent 2

D94592	1 x 3 ml	Creatinine Standard	
D98485	5 x 3 ml	Calibrator	Diacal Auto
D98481	12 x 5 ml	Control normal	Diacon N
D98482	12 x 5 ml	Control abnormal	Diacon P

1. Reagent and Sample Preparation:

The reagents are ready to use.
Urine: for urine preparation please refer to packing insert.

2. Instrument settings:

Temperature: 37 °C

GENERAL	Test name	Creatinine
	Analysis mode	Fixed Time bi-reagent
	Sample type	Serum
	Units	mg/dL
	Reaction type	Ascending
	Decimals	2
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Monochromatic
Volumes	Sample	10
	Reagent 1	200
	Reagent 2	50
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	505
	Reference	-
Times	Reading 1	240
	Reading 2	360
	Reagent 2	120
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	3
	Calibration curve	-
OPTIONS	Blank absorbance limit	0,500
	Kinetic blank limit	-
	Linearity limit	15

NOTE:
These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.

Компания Трион:
Официальный сайт - www.trionmed.ru
Продажа домашней медтехники в розницу - www.trionmed.com
Лабораторное оборудование Dialab - www.kuban-pm.ru