

BIOSYSTEMS A25
**CHOLESTEROL HDL, DIRECT
IMMUNOINHIBITION**

Liquid
2 Reagents

REF

Cont.

F03100	5 x 100 ml	4 x 100 ml	Reagent 1
		1 x 100 ml	Reagent 2
F03115	5 x 50 ml	4 x 50 ml	Reagent 1
		1 x 50 ml	Reagent 2

Additionally offered:

F03710SV	1 x 3 ml	HDL Cholesterol Calibrator
D99486	3 x 3 ml	Lipid Control normal Diacon Lipids

1. Reagent preparation

The reagents are ready to use.

2. Instrument setting:

Temperature: 37°C

GENERAL	Test name	HDL – Cholesterol
	Analysis mode	2 Reagents Differential
	Sample type	Serum
	Units	mg/dL
	Reaction type	Ascending
	Decimals	0
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Bichromatic
Volumes	Sample	3
	Reagent 1	300
	Reagent 2	-
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	600
	Reference	-
Times	Reading 1	300
	Reading 2	600
	Reagent 2	300
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	3
	Calibration curve	-
OPTIONS	Blank absorbance limit	0,05
	Kinetic blank limit	-
	Linearity limit	180

NOTE:

These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.

:

-

Dialab -

www.trionmed.ru
- www.trionmed.com
www.kuban-pm.ru