APPLICATION PROCEDURE

BIOSYSTEMS A25

LDH-L

IFCC

Liquid
2 Reagents

REF	Cont.		
D00663	**5 x 50 ml**	4 x 50 ml	Reagent 1
		1 x 50 ml	Reagent 2

D98485	5 x 3 ml	Calibrator	Diacal Auto
D98481	12 x 5 ml	Control normal	Diacon N
D98482	12 x 5 ml	Control abnormal	Diacon P

1. Reagent preparation:

The reagents are ready to use.

2. Instrument settings:

Temperature: 37°C

GENERAL	Test name	LDH-L
	Analysis mode	Kinetic
	Sample type	Serum
	Units	U/L
	Reaction type	Ascending
	Decimals	0
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Monochromatic
Volumes	Sample	4
	Reagent 1	200
	Reagent 2	50
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	340
	Reference	
Times	Reading 1	180
	Reading 2	360
	Reagent 2	120
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	1
	Calibration curve	-
OPTIONS	Blank absorbance limit	1,400
	Kinetic blank limit	-
	Linearity limit	1512

NOTE:
These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.

Компания Трион:
Официальный сайт - www.trionmed.ru
Продажа домашней медтехники в розницу - www.trionmed.com
Лабораторное оборудование Dialab - www.kuban-pm.ru