

Α ΕΛΛΗΝΙΚΗ ΔΙΑΓΝΩΣΤΙΚΗ ΚΑΙ ΕΡΕΥΝΑ

RHEUMATOID FACTOR (RF)

Diagnostic reagent for the quantitative in vitro determination of Rheumatoid Factor (RF) in human serum by turbidimetric assay

REF	Cont.
A02530	1 x 25 ml RF Antibody Reagent 5 x 25 ml RF Buffer
A00723	1 x 1 ml RF Calibrator Super High
A00724	1 x 5 ml RF Calibrator Super High
A00810	1 x 1 ml RF Control
A00811	1 x 5 ml RF Control

1. Reagent preparation

Sample: Ready for use
 R1: Ready for use
 R2: Ready for use
 Calibration: Dilute the RF Calibrator high successively 1:2 in NaCl 9 g/L to set up a calibration curve. Use NaCl 9 g/L as zero point.

2. Instrument setting

Test	RFAB
Unit	IU/mL
Decimals	1
Analysis type	Two point end assay 1
Main wavelength	340
Sub wave length	
Method	Turbidimetry
Correction	1.00x + 0
Calibration type	Spline
Standard	Multipoint calibration
Normal range	0 – 50

Sample	12 µL
Automatic dilution	NO
R1	200 µL
R2	40 µL
Read Main	54
Sub	31
Prozone check	NO

250 Tests/Kit

: -

Dialab - www.trionmed.ru
www.trionmed.com
www.kuban-pm.ru

Note: These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing tests.