

BIOSYSTEMS A25
TRIGLYCERIDES
GPO-PAP

Liquid
Single Reagent

REF	Cont.		
D98386B	1 x 1000 ml	Single Reagent	
D00389	5 x 100 ml	Single Reagent	
D96388	5 x 50 ml	Single Reagent	
D95380	1 x 3 ml	Triglycerides Standard	
D98485	5 x 3 ml	Calibrator	Diacal Auto
D99486	3 x 3 ml	Lipid Control normal	Diacon Lipids
D98481	12 x 5 ml	Control normal	Diacon N
D98482	12 x 5 ml	Control abnormal	Diacon P

1. Reagent preparation:

The reagent is ready to use.

2. Instrument settings:

Temperature: 37 °C

GENERAL	Test name	Triglycerides
	Analysis mode	Endpoint
	Sample type	Serum
	Units	mg/dL
	Reaction type	Ascending
	Decimals	0
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Bichromatic
Volumes	Sample	3
	Reagent 1	300
	Reagent 2	-
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	505
	Reference	670
Times	Reading 1	600
	Reading 2	-
	Reagent 2	-
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	3
	Calibration curve	-
OPTIONS	Blank absorbance limit	0,150
	Kinetic blank limit	-
	Linearity limit	1300

NOTE:

These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.

:

-